


ИНФРАКРАСНЫЙ ТЕРМОМЕТР T1


Особенности:

- Прибор специально разработан для быстрого и точного измерения температуры тела человека на лбу;
- Показание измерения отображается через секунду;
- ЖК-дисплей с подсветкой;
- Автоматическое отключение спустя 60 с простоя;
- Сохранение в памяти до 32 наборов данных;
- Активируемая функция звуковой сигнализации;
- Наличие сертификата CE.


Технические характеристики

Материал	АБС-пластик
Цвет	Белый / серый
Размеры (ДхШхВ)	(105 x 48 x 150) мм
Масса	165 г
Разрешение дисплея	0,1 °C
Диапазон измерения	+32 ... +42,9 °C
Расстояние измерения	Не больше 5 см
Погрешность измерения	В диапазоне +35 ... +42 °C: ±0,2 °C Вне диапазона +35 ... +42 °C: ±0,3 °C
Температура эксплуатации	+16 ... +35 °C
Температура хранения	-20 ... +55 °C
Относительная влажность	≤ 10-93%
Источник питания	2 батареи 1,5 В АА
Гарантия	12 месяцев
Стандарт	ISO 13485, CE





Attestation of Conformity

No. ICR Polska/M6330215 

Name and address of Registered Manufacturer: Shenzhen Longin Technology Co., Ltd
Room 503, Haolong Commercial Building, No.138 Shajing Section, Guangshen Road, Shangye Community, Xinqiao Street, Baoan District, Shenzhen

Product name: Infrared forehead thermometer

Product type/model: T1

Trade mark: n/a

This Attestation confirms that the product meets the requirements of the following normative documents and within limits of its documents gives presumption of conformity with essential requirements of Directive 93/42/EEC.

Relevant EC Directive: Medical Device Directive 93/42/EEC

Conformity assessment procedure: EC Declaration of Conformity (Annex VII of Directive 93/42/EEC)

Classification: Class I according Rule 1 of Annex IX of Directive 93/42/EEC

Applied normative documents: EN 60601-1:2006+A1:2013+A12:2014, EN 60601-1-1:2015, EN 60601-1-2: 2015, EN 55011: 2016/A1: 2017

Applied Quality Management System

This AoC will remain valid only if Quality Management System Certificate remains valid. The assessment process has been carried out in accordance with the program PC-P-07-07. Evaluation has been carried out in accordance with test report made by: United Testing Technology (Hong Kong) Limited


No. of test report: A20032886SR-01, A20032886ER-01

Issue date: 25.03.2020

Expiration date: 24.03.2025

The mutual obligations and rights of the certification are regulated by the contract No. ICR Polska/2020-3125.


This Attestation applies to products having the same attributes (parameters), intended use, that have been evaluated and meet the requirements of the aforementioned standard.



Director: Rafal Kalinowski

Warsaw, 25.03.2020.

ICR Polska Co. Ltd.
ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com, e-mail: icrpolska@icrga.com



БУДЬТЕ ЗДОРОВЫ!